Print on Plain White Paper (Part No. 122919)

MIDMARK® Installation Instructions

9A101002 Colposcope Mounting Bracket

KIT COMPONENTS

Part Number	Description	Qty.
9A101002	Colposcope Mtg. Bracket Kit	1
003-0431-00	Installation Instructions	1

REQUIRED TOOLS

7/16” Wrench or Socket
5/32" & 3/16" Allen Wrench
Power Hand Drill
9/32” Drill Bit

IMPORTANT INSTRUCTIONS

Provides a means to hold a Zeiss Colposcope on examination table.

INSTALLATION (1)



DANGER
Unplug chair from electrical outlet.

INSTALLATION (2)



A. Remove two top drawers on foot end of table.
B. Pull footrest out to full extension.
C. Raise and support seat section.

INSTALLATION (3)



NOTE
Locate mounting plate so top edge is even with top edge of upper wrap. Locate large hole in plate over spot weld that is near seat support on upper wrap.

A. Hold mounting plate on upper wrap and mark top two holes then remove mounting plate.
B. Drill two 9/32” holes at location marks.
C. Install mounting plate using two 1/4”-20 x 5/8” screws from inside of upper wrap into top holes of plate.

© Midmark Corporation - 1998

INSTALLATION (4)



NOTE
Install support weldments so arms are angled toward foot end of table.

- **A.** Install upper support weldment using four 1/4”-20 x 5/8” screws.
- **B.** Install mounting post to lower support using two 1/4”-20 x 1 1/2” screws.
- **C.** Install mounting post to upper support with two 1/4-20 x 1 1/2” screws.

INSTALLATION (5)



- **A.** Using a level, assure mounting post is straight vertically.
- **B.** Mark location of four mounting holes of lower support weldment on cabinet side panel.
- **C.** Remove lower support weldment and mounting post from upper support weldment.

INSTALLATION (6)



- **A.** Drill four 9/32” holes at location marks.
- **B.** Install lower support weldment and mounting post and secure to cabinet side panel with four 1/4”-20 x 5/8” screws and nuts.

INSTALLATION (7)



- **A.** Push footrest in to stored position.
- **B.** Install two top drawers.
- **C.** Plug in table.

MIDMARK®
Midmark Corporation, 60 Vista Drive, P.O. Box 286, Versailles, Ohio 45380-0286
Phone: 1-937-526-3662 Fax: 1-877-249-1793 midmark.com

003-0431-00 Rev. B (7/98)

© Midmark Corporation - 1998